

Code	Range	Graduation	Accuracy	Hysteresis	Dial reading
2896-05 *	0.5mm	0.01mm	13µm	5µm	0-25-0

*Supplied with manufacturer inspection certificate



34.6