

COATING THICKNESS GAUGE (ADVANCED TYPE) CODE 5402-TC21

- Magnetic induction probe (FM) measures the thickness of non-magnetic coating and non-metallic coating on magnetic metal substrate.
 Substrate: steel, iron, alloy, hard magnetic steel, etc.
 Coating: zinc, aluminum, chrome, copper, rubber, paint, etc.
- Eddy current probe (NM) measures the thickness of non-conductive coating on non-magnetic metal substrate.
 - Substrate: copper, aluminum, zinc, tin, etc.
 - Coating: rubber, paint, plastic, anodized film, etc.
- Real-time temperature compensation guarantees high accuracy, thin plating and oxide layer less than 20µm can be measured accurately
- Reduces the effects of electromagnetic interference and hand-held operation
- Probe can be re-matched after abrasion
- Tolerance measurement with adjustable alarm threshold
- USB and bluetooth interface for data transmission and online measurement in real-time
- Coupling status indication
- Support cable printer



ONLINE MEASUREMENT

BLUETOOTH



mid-range magnetic induction probe FL (OPTIONAL)



high-range magnetic induction probe FX (OPTIONAL)



low-range magnetic induction probe FS (OPTIONAL)



TEMPERATURE

high-temp magnetic induction probe FH (OPTIONAL)



eddy current probe NM (OPTIONAL)

SPECIFICATION

Probe		FM (included) magnetic induction probe	FL (optional) mid-range magnetic induction probe	FX (optional) high-range magnetic induction probe	FS (optional) low-range magnetic induction probe	FH (optional) high-temp magnetic induction probe	NM (optional) eddy current probe
Range		0~1500μm	0~3000µm	0~10000µm	0~500μm	0~3000µm	0~1500μm
Resolution		0.1µm (<100µm 1µm (100µm~10					
Accuracy *	zero calibration	±(1µm+2%L)	±(1µm+3%L)	±(2µm+5%L)	±(1µm+2%L)	±(1µm+3%L)	±(1µm+2%L)
Accuracy	multi-point calibration	±(1µm+1%L)	±(1µm+2%L)	±(1µm+3%L)	±(1µm+1%L)	±(1µm+2%L)	±(1µm+1%L)
Measuring	mode	single point mea	asurement, scan r	node, differentia l r	mode, average mo	ode	
Calibration	mode	zero calibration,	one-point calibra	tion, two - point ca l	ibration, mu l ti <mark>-</mark> poir	nt ca l ibration	
Minimum s	ubstrate thickness	0.5mm	0.5mm	2mm	0.2mm	0.5mm	0.3mm
Minimum n	neasuring area	Ø7mm	Ø7mm	Ø40mm	Ø3mm	Ø7mm	Ø5mm
Minimum c	urvature radius of rkpiece	1.5mm	1.5mm	10mm	1mm	1.5mm	3mm
Data storaç	je	500 groups					
Interface		USB, bluetooth					
Operation t	emperature	-10°C~50°C					
Power sup	ply	3×1.5V AAA bat	teries				
Dimension		150×70×30mm					
Weight		160g					

^{*} L is measuring thickness in µm

STANDARD DELIVERY

Cable printer

Main unit	1 pc
Magnetic induction probe (FM)	1 pc
Zero calibration block for FM probe	1 pc
Calibration foils (12/50/100/250/500/1000μm)	6 pcs
AAA battery	3 pcs
Software and USB cable	1 pc
OPTIONAL ACCESSORY	1 pc
	1 pc 5401-TC11-NM
OPTIONAL ACCESSORY Eddy current probe (NM)	
OPTIONAL ACCESSORY Eddy current probe (NM) (with zero calibration block for NM probe)	5401-TC11-NM
OPTIONAL ACCESSORY Eddy current probe (NM) (with zero calibration block for NM probe) Mid-range magnetic induction probe (FL)	5401-TC11-NM 5402-TC21-FL

5401-TC11-PRINTER