

BACK PLUNGER TYPE DIAL INDICATOR



2320-02

handle can be removed



- Supplied with handle
- Optional accessory: contact points

Part No.	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading
2320-02*	.2"	.001"	±.001"	.00033"	.1"	0-100

* Supplied with manufacturer inspection certificate traceable to NIST USA

points (included)

