

DIAL INDICATORS

INSIZE^{PLUS}

- Meet DIN878
- With limit pointers
- Optional accessory: backs, contact points



2892-10

**INSPECTION
CERTIFICATE**

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Maximum measuring force	Cap	Remark
2892-10 *	10mm	0.01mm	17μm	3μm	1mm	0-100	1.5N	can lift	lug back
2892-10F *	10mm	0.01mm	17μm	3μm	1mm	0-100	1.5N	can lift	flat back

* Supplied with manufacturer inspection certificate

